

Data Table

Test Device	Contact Time	Cycle	Test Microorganism	
			<i>S. epidermidis</i> ATCC 12228	<i>E. coli</i> ATCC 11229
Strava Splint Bath	Time Zero	Maintenance Cycle/UV	76,666	1,033,333
	1 Minute		163	895
	10 Minutes		0	0
	30 Minutes		0	0
	60 Minutes		0	0
Test Device	Contact Time	Cycle	Test Microorganism	
			MS2 ATCC 15597-B1	<i>M. terrae</i> ATCC 15755
Strava Splint Bath	Time Zero	Sanitation Mode	62,333,333	606,666
	2 Minutes		0	0
	5 Minutes		0	0
	10 Minutes		0	0